SERIES 222 DIAL INDICATOR

Certificate Jeweled

- Graduation: 0.01mm
- With limit hands
- Jeweled bearing
- Range per revolution: 1mm
- With lock screw



222-010-11


Styli(optional)



Series 270

Application





Code	Range	Back	Accuracy	Hysteresis	Scale
222-003-11	3mm	lug	14μm	3μm	0-100
222-005-11	5mm	lug	16μm	3μm	0-100
222-010-11	10mm	lug	17μm	3μm	0-100
222-003-12	3mm	flat	14μm	3μm	0-100
222-005-12	5mm	flat	16μm	3μm	0-100
222-010-12	10mm	flat	17μm	3μm	0-100

With manufacturer inspection certificate

SERIES 222 DIAL INDICATOR


Certificate Jeweled

- Graduation: 0.001mm
- With limit hands
- Jeweled bearing
- Rang per revolution: 0.2mm
- With lock screw



222-001-01

unit: mm



Styli(optional)



Series 270

Application



Code	Range	Back	Scale	Accuracy	Hysteresis
222-001-01	1mm	lug	0-100-0	5μm	2μm
222-001-02	1mm	flat	0-100-0	5μm	2μm

With manufacturer inspection certificate