SERIES 223 DIAL INDICATOR

Certificate

- Graduation: 0.01mm
- With limit hands
- Range per revolution: 1mm
- With lock screw



223-010-11

unit: mm



Styli(optional)



Application



Code	Range	Back	Accuracy	Hysteresis	Scale
223-010-11	10mm	lug	20μm	3μm	0-100
223-010-12	10mm	flat	20μm	3μm	0-100

With manufacturer inspection certificate