228 DIAL INDICATOR


Certificate Jeweled

- Bezel: Ø40mm
- Graduation: 0.001mm
- With limit hands
- Jeweled bearing
- Supplied with lug and flat back



228-001-01

unit: mm



Application



Contact points **(optional)**



Series **270**

Code	Range	Accuracy	Hysteresis	Range/rev	Scale
228-001-01	1mm	5μm	2μm	0.1mm	0-100

With manufacturer inspection certificate


SERIES 228 DIAL INDICATOR

- Bezel: Ø40mm
- Graduation: 0.01mm
- With limit hands
- Jeweled bearing
- Rang per revolution: 1mm
- Supplied with lug and flat back



228-005-11

unit: mm



Styli(optional)



Application



Code	Range	Accuracy	Hysteresis	Scale	b	c
228-003-11	3mm	14μm	3μm	0-100	11	44
228-005-11	5mm	16μm	3μm	0-100	12.5	45.5

With manufacturer inspection certificate