SERIES 263 LONG STYLI DIAL TEST INDICATOR

Certificate Jeweled

- Jeweled bearing
- Double scale spacing
- Anti-magnetic copper body
- With two clamps: Ø4mm and Ø8mm



263-008-11L

Application



Clamps **(included)**





unit: mm



Code	Range	Bezel(d)	Graduation	Styli	Accuracy	Hysteresis	Scale	L
263-008-11	0.8mm	Ø37.5mm	0.01mm	carbide	13μm	3μm	0-40-0	30.5
263-008-11L	0.8mm	Ø37.5mm	0.01mm	ruby	13μm	9μm	0-40-0	46.4

With manufacturer inspection certificate