

Muayene Sertifikalı Seramik Gösterge Bloğu Setleri Sınıf 0

Kontrol sertifikası Sınıf 0



Kontrol
sertifikası

Metrik

No.	Set başına bloklar	Derece	Malzeme	Set içeriği	Ağırlık [g]
516-338-10	112	0	Seramik	1,0005 mm (1 blok) 1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-24,5 mm (0,5 mm adım / 49 blok) 25-100 mm (25 mm adım / 4 blok)	3200
516-342-10	103	0	Seramik	1,005 mm (1 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-24,5 mm (0,5 mm adım / 49 blok) 25-100 mm (25 mm adım / 4 blok)	3000
516-370-10	88	0	Seramik	1,0005 mm (1 blok) 1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-9,5 mm (0,5 mm adım / 19 blok) 10-100 mm (10 mm adım / 10 blok)	3200
516-346-10	87	0	Seramik	1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-9,5 mm (0,5 mm adım / 19 blok) 10-100 mm (10 mm adım / 10 blok)	2800
516-350-10	76	0	Seramik	1,005 mm (1 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-9,5 mm (0,5 mm adım / 19 blok) 10-40 mm (10 mm adım / 4 blok) 50-100 mm (25 mm adım / 3 blok)	2400
516-354-10	56	0	Seramik	0,5 mm (1 blok) 1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-24 mm (1 mm adım / 24 blok) 25-100 mm (25 mm adım / 4 blok)	2200
516-358-10	47	0	Seramik	1,005 mm (1 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-24 mm (1 mm adım / 24 blok) 25-100 mm (25 mm adım / 4 blok)	2000
516-362-10	47	0	Seramik	1,005 mm (1 blok) 1,01-1,09 mm (0,01 mm adım / 19 blok) 1,2-1,9 mm (0,1 mm adım / 8 blok) 1-9 mm (1 mm adım / 9 blok) 10-100 mm (10 mm adım / 10 blok)	2600
516-395-10	46	0	Seramik	1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-9 mm (1 mm adım / 9 blok) 10-100 mm (10 mm adım / 10 blok)	2700
516-179-10	34	0	Seramik	1,0005 mm (1 blok) 1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-5 mm (1 mm adım / 5 blok) 10 mm (1 blok)	1200
516-366-10	32	0	Seramik	1,005 mm (1 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-9 mm (1 mm adım / 9 blok) 10-30 mm (10 mm adım / 3 blok) 60 mm (1 blok)	1300

Özellikler

Hata Sınırı

EN ISO 3650

Mitutoyo Certificate number: T00E10309 page 1 of 1

CERTIFICATE OF INSPECTION

Instrument: Gauge Block
Grade: 0 (ASME)
Date of inspection: 18th May, 2006 Unit: µinch

Number	Length inch	Ident. No.	Central Deviation	Maximum Deviation	Minimum Deviation	Variation
1	0.0230		-0.4	+1.2	-0.4	1.6

ENVIRONMENT: Air temperature: (50 ± 1.0)°F (10 ± 0.5)°C
BASIS OF TEST: ASME B89.1.9-2002
MATERIAL: Steel
INSPECTION METHOD: by comparison
COEFFICIENT OF THERMAL EXPANSION: (6.0 ± 0.3) × 10⁻⁶/°F [(10.8 ± 0.5) × 10⁻⁶/°C]
EXPANDED UNCERTAINTY: ± 0.0001 inch or less 1.4 µinch
Per Deviation Deviation: Up to 0.001 inch (0.025 mm) or less
(n=2)
QUALITY SYSTEM: ISO 9001 Registered Firm / JQA NUMBER
TRACEABILITY: Traceable to NIST No. 601/50004-03
0103-010001 Certificate of Standard and Technology

Date: 18th May, 2006
A. Shiro

MITUTOYO CO., LTD.
1-1-1, Higashi-Shinjuku, Shinjuku-Ku, Tokyo 163-8601, Japan
Tel: 03-3358-0001 Fax: 03-3358-0002

Kontrol Sertifikası

Grad K/0/1/2 gösterge blok setleri JCSS kalibrasyon sertifikası (Mitutoyo JCSS akredite laboratuvarı tarafından verilmiştir) çevrimiçi bkz.