Muayene Sertifikalı Seramik Mastar Blok Setleri Sınıf 2

Kontrol sertifikalı Sınıf 2

Kontrol sertifikası

Metrik

No.	Set başına bloklar	Derece	Malzeme	Set İçeriği	Ağırlık [g]
516-340-10	112	2	Seramik	1,0005 mm (1 blok) 1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-24,5 mm (0,5 mm adım / 49 blok) 25-100 mm (25 mm adım / 4 blok)	3200
516-344-10	103	2	Seramik	1,005 mm (1 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-24,5 mm (0,5 mm adım / 49 blok) 25-100 mm (25 mm adım / 4 blok)	3000
516-372-10	88	2	Seramik	1,0005 mm (1 blok) 1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-9,5 mm (0,5 mm adım / 19 blok) 10-100 mm (10 mm adım / 10 blok)	3200
516-348-10	87	2	Seramik	1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-9,5 mm (0,5) mm adım / 19 blok) 10-100 mm (10 mm adım / 10 blok)	2800
516-352-10	76	2	Seramik	1,005 mm (1 blok) 1,01-1,49 mm (0,01 mm adım / 49 blok) 0,5-9,5 mm (0,5 mm adım / 19 blok) 10-40 mm (10 mm adım / 4 blok) 50-100 mm (25 mm adım / 3 blok)	2400
516-356-10	56	2	Seramik	0,5 mm (1 blok) 1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-24 mm (1 mm adım / 24 blok) 25-100 mm (25 mm adım / 4 blok)	2200
516-360-10	47	2	Seramik	1,005 mm (1 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-24 mm (1 mm adım / 24 blok) 25-100 mm (25 mm adım / 4 blok)	2000
516-364-10	47	2	Seramik	1,005 mm (1 blok) 1,01-1,09 mm (0,01 mm adım / 19 blok) 1,2-1,9 mm (0,1 mm adım / 8 blok) 1-9 mm (1 mm adım / 9 blok) 10-100 mm (10 mm adım / 10 blok)	2600
516-397-10	46	2	Seramik	1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-9 mm (1 mm adım / 9 blok) 10-100 mm (10 mm adım / 10 blok)	2700
516-181-10	34	2	Seramik	1,0005 mm (1 blok) 1,001-1,009 mm (0,001 mm adım / 9 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-5 mm (1 mm adım / 5 blok) 10 mm (1 blok)	1200
516-368-10	32	2	Seramik	1,005 mm (1 blok) 1,01-1,09 mm (0,01 mm adım / 9 blok) 1,1-1,9 mm (0,1 mm adım / 9 blok) 1-9 mm (1 mm adım / 9 blok) 10-30 mm (10 mm adım / 3 blok) 60 mm (1 blok)	1300

Özellikler

Hata Sınırı	EN ISO 3650

Mitutoyo

Certificate number T06E10369
page 1 of 1

CERTIFICATE OF INSPECTION

Instrument: Gauge Block
Grade: 0 (ASME)
Date of inspection: 18th May. 2006

Unit : μ inch

Nominal Length inch	Ident. No.	Central Deviation	Maximum Deviation	Minimum Deviation	Variation
1	06280	-0.4	+1.2	-0.4	1.6

ENVIRONMENT: Air temperature (68 ± 1.8)F [(20 ± 1.0) ℃]
BASIS OF TEST: ASME B89.1.9-2002
MATERIAL: Steel
INSPECTION METHOD: by comparator
COEFFICIENT OF THERMAL EXPANSION: (6.0±0.3)×10-6/F [(10.8±0.5)×10-6/K]
EXPANDED UNCERTAINTY (For Central Deviation (k=2)): 4" or less 2.4 μ inch; Up to20" (1.2 + L / 2.3) μ inch (L=Nominal length) L:inch
QUALITY SYSTEM: ISO 9001 Registered Firm : JQA JM0189
TRACEABILTY: Traceable to NIST No.821/266834-03 (NIST=National Institute of Standard and Technology)

Date 18th May. 2006

S. Sterab

Kontrol Sertifikası

Grad K/0/1/2 gösterge blok setleri JCSS kalibrasyon sertifikası (Mitutoyo JCSS akredite laboratuvarı tarafından verilmiştir) çevrimiçi bkz.

Tüm ürünler ticari müşterilere satılır.
Ürün resimleri herhangi bir yükümlülük getirmez. Ürün açıklamaları, özellikle de tüm teknik özellikler, yalnızca üzerinde açıkça mutabakata varıldığında bağlayıcıdır.