SERIES 224 DIAL INDICATOR


Certificate

Jeweled

- Graduation: 0.01mm
- With limit hands
- Jeweled bearing
- Range per revolution: 1mm
- With lock screw



224-010-11



Styli(optional)



Series 270

Application



Package



Code	Range	Back	Accuracy	Hysteresis	Scale
224-010-11	10mm	lug	17μm	3μm	0-100
224-010-12	10mm	flat	17μm	3μm	0-100

With manufacturer inspection certificate