SERIES 228 COMPACT DIAL INDICATOR


Certificate Jeweled

- Meet DIN878
- Graduation: 0.01mm
- With limit hands
- Jeweled bearing
- With lock screw



228-005-61P

unit: mm



Styli(optional)



Application



Code	Range	Accuracy	Hysteresis	Range/rev	Scale	Back	b	c
228-003-61P	3mm	12μm	3μm	0.5mm	0-50	lug	10.5	46
228-005-61P	5mm	14μm	3μm	0.5mm	0-50	lug	12.5	48
228-003-62P	3mm	12μm	3μm	0.5mm	0-50	flat	10.5	46
228-005-62P	5mm	14μm	3μm	0.5mm	0-50	flat	12.5	48

With manufacturer inspection certificate