261 PRECISION DIAL TEST INDICATOR


Certificate

Jeweled


- Meet DIN2270
- Jeweled bearing
- Double scale spacing
- With two clamps: Ø6mm and Ø8mm




Clamps **(included)**



unit: mm





Code	Range	Bezel(d)	Graduation	Styli	Accuracy	Hysteresis	Scale
261-014-01	0.14mm	Ø38mm	0.001mm	ruby	4μm	2μm	0-70-0

With manufacturer inspection certificate

Application

