JEWELED BEARING

INSPECTION CERTIFICATE

PRECISION DIAL INDICATORS



2313-1A

- Graduation: 0.001mm
- Jeweled bearing
- Supplied with limit pointers and bezel clamp
- Optional accessory: backs, contact points



Code	Range	Accuracy	Hysteresis	Range/rev	Dial reading	Maximum measuring force	Cap	Remark
2313-1FA*	1mm	5μm	2μm	0.2mm	0-100-0	2N	cannot lift	flat back
2313-2FA*	2mm	6μm	2.5μm	0.2mm	0-100-0	2N	cannot lift	flat back
2313-1A*	1mm	5μm	2μm	0.2mm	0-100-0	2N	cannot lift	lug back
2313-2A*	2mm	6μm	2.5μm	0.2mm	0-100-0	2N	cannot lift	lug back

*Supplied with manufacturer inspection certificate